

# **APC International, Ltd. Piezo $d_{33}$ Test System**



**APC International, Ltd**

**APC Part number: 90-2030.1**

**APC International, Ltd. All rights reserved**

## DESCRIPTIONS

The part number 90-2030.1 ( $d_{33}$  meter) is a special instrument for directly measuring the piezoelectric constant  $d_{33}$  values of piezoelectric ceramics and single crystals. This meter can also measure the  $d_{33}$  values in various single crystal directions for lithium niobate, quartz, and tourmaline. It is capable of measuring the  $d_{33}$  value over a very large range, at high resolution, and with a high degree of reliability. The measurement is quick and easily made with minimal training required. A variety of sizes and shapes can be easily accommodated and measured that includes discs, blocks, rings, tubes and semispherical shells. The direct value readout is displayed on a 3 ½ inch digital meter. This instrument is invaluable as a tool for quality assurance, production in-line inspection and research applications.

## SPECIFICATIONS

$D_{33}$  Range: x1 range: 10 to 2000 pC/N  
x 0.1 range: 1 to 200 pC/N

Accuracy: x 1 range:  $\pm 2\%$  of the  $d_{33}$  value in 100 to 2000 pC/N  
 $\pm 5\%$  of the  $d_{33}$  value in 10 to 200 pC/N  
x 0.1 range:  $\pm 2\%$  of the  $d_{33}$  value in 10 to 200 pC/N  
 $\pm 5\%$  of the  $d_{33}$  value in 1 to 20 pC/N

Resolution: x 1 range: 1 pC/N  
x 0.1 range 0.1pC/N

Force frequency: 110Hz Amplitude 0.25N

Polarity indication:

Indicates polarity on upper face of test element in compression.  
(- sign means negative, no sign is positive)

Shunt Capacitance; 1pF (for x 1 range)  
0.1pF (for x 0.1 range)

Dimensions: Force Head: 110x140mm Chassis: 280x200x90mm

Weight: Force Head: 3kg Chassis: 2kg





## FRONT PANEL OVERVIEW

1. **D<sub>33</sub> Range Key:** Choose from x1 or x0.1. Pushing down (x 1), the measuring range is from 10 to 2000 pC/N, resolution is 1 pC/N for normal piezo material such as piezo ceramics. Pushing up (x0.1), the measuring range is 1 to 100 pC/N, resolution is 0.1 pC/N for low piezo constant material such as piezo crystal.
2. **Force and D<sub>33</sub> Key:** When pushing down the force light is on. The digital display shows the force put on the sample by the shaker. The display should show at  $(250 \pm 10) \times 10^{-3}$  N. If the force is too high or low  $(250 \pm 20) \times 10^{-3}$  N, you must turn the knob of the shaker screw for adjusting the clamping force on the test sample.
3. The 3 ½" digital display shows the force put on the test sample of the d<sub>33</sub> value of the test sample. The display indicates polarity on the upper face of the test element in compression. (- sign means negative and no sign is positive).
4. The power switch turns off the display and the display light of the force or d<sub>33</sub> will turn on.
5. **Zero adjust screw:** adjusting this screw can get system balance for a high accuracy measuring result. (see calibration)



## **BACK PANEL OVERVIEW**

1. Output plug: A three-pin plug is connected with the shaker to the supply power and with an output of 110Hz and 0.25 N force.
2. Input plug: An eight-pin plug is connected with the shaker and picks up the signal from the shaker.
3. AC power plug: Universal AC 60/50 Hz, 110-240 V AC plug with a spare 2-amp fuse.



## SHAKER OVERVIEW

1. Adjust knob: Turn the adjusting knob to drive the top probe up or down. The test sample will be clamped between the top probe and the bottom probe. Do not screw the sample too tight.
2. The top and bottom probe transfers a signal from the test sample to the measuring system and holds the sample driving the sample vibration.
3. Output and input connector: There is a three-pin plug that connects the meter to the shaker and an eight-point output plug that sends the signal to the meter.

## SETUP AND CALIBRATION

1. Connect the meter input and output plug with the shaker input and output plug by the two cables supplied with the system.
2. Turn the meter on for 15 minutes and then put the standard test sample between the top and bottom probes.
3. Turn the Force and  $D_{33}$  switch to Force. The meter read out should be  $(250 \pm 10) \times 10^{-3} \text{ N}$ . If it isn't, adjust the probes knob until the meter read out is  $(250 \pm 10) \times 10^{-3} \text{ N}$ .
4. Turn the Force and  $D_{33}$  switch to  $D_{33}$ . The meter read out should be close to the  $D_{33}$  value of the standard test sample. Write down the value of the meter displayed for D1. Turn the test sample 180 degrees and measure the sample again. The meter will show the opposite polarization and the  $D_{33}$  value for D2. If the two values are different (D1 & D2) then calculate the D3 by  $(D1 + D2) \times 0.5$ . Using the Zero point adjust knob get the



meter display to show the D3 value. Now the test sample negative and positive should be the same.

5. Adjust the back panel calibration screw and let the meter display show the same value as the standard test marked.
6. The manufacturer calibrates the meter before it is shipped but recalibrating to begin using and once a month is suggested.

## OPERATION

1. Put the test parts between the two probes. Clamp the parts as close to the center as possible. Turn the probes adjust knob lightly and clamp the test parts. Do not over tighten. When the meter display is stable it will display the results. Parts that are clamped too tight or too loose will impact the test results.
2. If you want to measure same thickness parts, you can push lightly down on the bottom probe and then change the test parts. There is no need to adjust the knob.
3. When measuring large capacitance samples, the meter read out value needs modified or the accuracy will be more than 1%. The following equation can be used to calculate the true value.

$$d_{33} \text{ (modified)} = d_{33} \text{ (Display)} \times (1+Cc) \text{ for switch on "x1"}$$

$$d_{33} \text{ (modified)} = d_{33} \text{ (Display)} \times (10+Cc) \text{ for switch on "x0.1"}$$

Cc= capacitor of sample ( $\mu$ F)

## CALCULATE E33 AND G33

The relative dielectric constant  $E_{33}=(TxC)\div(E0xS)$

T= The thickness of the sample (m)

S= The area of the sample electrode ( $m^2$ )

C= The capacitance of the sample (F)

$E0= 8.85 \times 10^{-12}$  F/m

The piezo voltage  $g_{33}=d_{33}\div E_{33}$

## MAINTENANCE

This system is maintenance free. If you experience any problems please contact us for assistance.

## ACCESSORIES

- |                                       |          |
|---------------------------------------|----------|
| 1. Three wire cables with connectors  | 1 piece  |
| 2. Seven wire cable with connectors   | 1 piece  |
| 3. Power cord                         | 1 piece  |
| 4. Cone shaped probe head             | 2 pieces |
| 5. Standard test sample (calibration) | 1 piece  |
| 6. Fuse- 2amp                         | 1 piece  |
| 7. Operating manual                   | 1 piece  |